

ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

SV-0511M

System: GNDRAD PIR

Site: Nauru

OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Date of Change (UTC)	Time Start UTC	Time End UTC
PIR	31304F3	Gndrad	11/04/2004	-10.28 K0 .28169 K1 1.018 K2 -2.797 K3	na	11/15/2005	03:00	04:15

NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Change Logger Coefficients and Version
PIR	31303F3	Gndrad	11/15/2005	-5.11 K0 .28951 K1 1.000 K2 (-3.060) K3	na	X

Replacement completed by: Troy Culgan

NOTES:

Old Part remains on site as spare:

Old part packed for shipment to: SGP

Other old part disposition (explain):

Other comments: